

United States Geological Survey

Reston Stable Isotope Laboratory

Report of Stable Isotopic Composition

Reference Material NBS 22-0.25 μL

(Hydrogen and Carbon Isotopes in Oil Sealed in a Silver Tube)

This reference material (RM) is intended for calibration of stable hydrogen ($\delta^2\text{H}$) and carbon ($\delta^{13}\text{C}$) measurements of unknown hydrogen- or carbon-bearing substances with a TC/EA (thermal conversion/elemental analyzer), an EA (elemental analyzer), and an isotope-ratio mass spectrometer by quantifying drift with time and isotope-ratio-scale contraction. This RM consists of 0.25 μL of NBS 22 oil [1] sealed in a silver tube [2]. This RM is issued in quantities of 50 sealed silver tubes per bottle. There is no limit on distribution. NBS 22 was prepared by S. Silverman, Chevron Oil Company, La Habra, California. Although stocks are low, a unit containing 1 mL of NBS 22 oil may be available from the International Atomic Energy Agency (IAEA) [3]. NBS 22 oil was available from the U.S. National Institute of Standards and Technology (NIST) [4] as RM 8539, but may be out of stock.

Recommended Values: Stable hydrogen isotopic compositions are expressed herein as delta values [5] relative to VSMOW (Standard Mean Ocean Water) on a scale normalized such that the $\delta^2\text{H}$ value of SLAP (Standard Light Antarctic Precipitation) is -428‰ , respectively [6,7]. Stable carbon isotopic compositions are expressed relative to VPDB (Vienna Peedee belemnite) on a scale normalized such that the $\delta^{13}\text{C}$ values of NBS 19 calcium carbonate and LSVEC lithium carbonate are $+1.95\text{‰}$ and -46.6‰ , respectively [8]. The isotopic compositions of NBS 22-0.25 μL are identical to the stable hydrogen isotopic composition [9] and the stable carbon isotopic composition [8] of NBS 22, except that each combined standard uncertainty value (μ_C) has been increased to account for hydrogen and carbon blanks, both of which were below detection.

Stable hydrogen isotopic composition: $\delta^2\text{H}_{\text{VSMOW-SLAP}} = -116.9 \pm 0.9\text{‰}$

Stable carbon isotopic composition: $\delta^{13}\text{C}_{\text{VPDB-LSVEC}} = -30.03 \pm 0.06\text{‰}$

Nominal volume of oil: 0.25 μL (Although the RSIL attempts to ensure that each silver tube has the same volume of oil, slight differences are observed owing to variations in the inside diameter of the silver tubing provided by the manufacturer. The typical relative variation in volume among 50 tubes is $\pm 3\%$, but this cannot be guaranteed.)

Technical coordination for this RM was provided by Haiping Qi of the RSIL.

Reston, Virginia 20192
September 22, 2011

Tyler B. Coplen, Director
Reston Stable Isotope Laboratory

Expiration of Reference Value: The reference value for the isotopic composition of NBS 22-0.25 μL is valid for a period of 10 years, provided the RM is handled in accordance with the instructions given in this Report of Stable Isotopic Composition (see “Instructions for Use”). The reference value is nullified if the RM is damaged by freezing or other means, contaminated, or otherwise modified.

Maintenance of RM Certification: The Reston Stable Isotope Laboratory (RSIL) will monitor this RM over the period of its certification. If substantive technical changes occur that affect the certification before the expiration of this report, the RSIL will notify the purchaser.

Distribution and Stability: NBS 22-0.25 μL is stable at normal room temperatures. The RSIL has monitored this RM for a period of two years since the reference water was sealed in the silver tubes and no change in isotopic composition has been observed. To minimize the potential for contamination, it is recommended that this RM be stored in the container in which it is supplied. The RM container should be sealed well after use to minimize tarnishing of the silver tubes.

Instructions for use: NBS 22 can be used at the beginning, the middle, and the end of the analysis sequence to enable satisfactory correction of drift with time. The amount of hydrogen or carbon in references and unknowns should be the same or similar to minimize bias in measurement results. Two or three silver tubes containing NBS 22 can be combined in a single port of a TC/EA carousel to increase the size of the sample.

Reporting of Stable-isotope-delta Values: The following recommendations are provided for reporting stable hydrogen and carbon isotope-delta values. It is recommended that:

- The $\delta^2\text{H}$ values of all hydrogen-bearing substances be expressed relative to VSMOW-SLAP on a scale where $\delta^2\text{H}_{\text{SLAP}2} = -427.5 \text{‰}$ or $\delta^2\text{H}_{\text{SLAP}} = -428 \text{‰}$ exactly [7,10].
- The $\delta^{13}\text{C}$ values of all carbon-bearing substances be expressed relative to VPDB on a scale such that the $\delta^{13}\text{C}$ values of NBS 19 calcium carbonate and LSVEC lithium carbonate are $+1.95 \text{‰}$ and -46.6‰ , respectively [8].
- Authors report δ values of international distributed (secondary) isotopic reference materials as though they had been interspersed among and used for normalization of unknowns, as appropriate, for the measurement method. In this manner, measurement results can be adjusted in the future as analytical methods improve and consensus values of internationally distributed isotopic reference materials change.
- Reporting of δ values relative to SMOW and PDB (Peedee belemnite) be discontinued [11].

REFERENCES

- [1] Silverman, S. R., 1964, Investigations of petroleum origin and evolution mechanisms by carbon isotope studies: In Craig H., Miller, S. L., Wasserburg, G. J. (eds), *Isotopic and Cosmic Chemistry*, North Holland, 92-102.
- [2] Qi, H., Gröning, M., Coplen, T. B., Buck, B., Mroczkowski, S. J., Brand, W. A., Geilmann, H., and Gehre, M., 2010, Novel silver-tubing method for quantitative introduction of water into high-

- temperature conversion systems for stable hydrogen and oxygen isotopic measurements: *Rapid Communications in Mass Spectrometry*, v. 24, p. 1821–1827.
- [3] International Atomic Energy Agency (IAEA), <http://www.iaea.org/programmes/aqcs/>
- [4] National Institute of Standards and Technology (NIST), <http://ts.nist.gov/measurementservices/referencematerials/index.cfm>
- [5] Coplen, T. B., 2011, Guidelines and recommended terms for expression of stable-isotope-ratio and gas-ratio measurement results: *Rapid Communications in Mass Spectrometry*, v. 25, 2538–2560.
- [6] Gonfiantini, R., 1978, Standards for stable isotope measurements in natural compounds: *Nature*, v. 271, p. 534–536.
- [7] Coplen, T. B., 1994, Reporting of stable hydrogen, carbon, and oxygen isotopic abundances: *Pure and Applied Chemistry*, v. 66, p. 273–276.
- [8] Coplen, T. B., Brand, W. A., Gehre, M., Gröning, M., Meijer, H. A. J., Toman, B., and Verkouteren, R. M., 2006, New guidelines for $\delta^{13}\text{C}$ measurements: *Analytical Chemistry*, v. 78, 2439–2441.
- [9] Coplen, T. B., and Qi, H., 2010, Applying the silver-tube introduction method for thermal conversion elemental analysis and a new $\delta^2\text{H}$ value for NBS 22 oil: *Rapid Communications in Mass Spectrometry*, v 24, 2269–2276.
- [10] International Atomic Energy Agency (IAEA), Reference Sheet for International Measurement Standards, http://nucleus.iaea.org/rpst/Documents/VSMOW2_SLAP2.pdf
- [11] Coplen, T. B., 1995, Discontinuance of SMOW and PDB: *Nature*, v. 375, 285.